

## Acid Test Mix on ZB-5MS

**Column:** Zebtron<sup>™</sup> ZB-5MS, GC Cap. Column 30 m x 0.25 mm x 0.25 µm, Ea

**Phase:** 5% Phenyl-Arylene 95% Dimethylpolysiloxane

**Dimensions:** 30 meters x 0.25 mm x 0.25 µm

**Order No:** 7HG-G010-11

**Oven Profile:** 50°C for 2 min to 250°C at 15°C/min for 5 min

**Carrier Gas:** Constant Flow Helium, 1.15 mL/min


**Injection:** Split :1 2 µL @ 250°C

**Detection:** Flame Ionization (FID) (325°C)

14951



Products used in this application:



### ANALYTES:

- 1 2,3-Butanediol
- 2 2-Ethylhexanoic acid
- 3 2,4-Dimethylphenol
- 4 Benzoic acid
- 5 Tridecane
- 6 Tetradecane
- 7 Pentadecane

