


**EPA Method 8082: Aroclor 1232 on ZB-5MSi****Column:** Zebron<sup>™</sup> ZB-5MSi, GC Cap. Column 30 m x 0.25 mm x 0.25  $\mu$ m, Ea**Phase:** 5% Phenyl 95% Dimethylpolysiloxane**Dimensions:** 30 meters x 0.25 mm x 0.25  $\mu$ m**Order No:** 7HG-G018-11**Oven Profile:** 120°C for 1min to 300 °C at 9 °C/min hold until last peak elutes**Carrier Gas:** Constant Flow Helium, 1.4 mL/min**Injection:** Split 70:1 1  $\mu$ L @ 210°C**Detection:** Electron Capture (ECD) (310°C)  
16416**Products used in this application:****ANALYTES:**

- 1 Tetrachloro-m-xylene
- 2 Decachlorobiphenyl

