

HPLC Application

ID No.: 19883

Reduction of RNA'ase Aeris WIDEPORE 3,6u C4 150x4.6mm

Column: Aeris[™] 3.6 μ m WIDEPORE C4 200 \AA , LC Column 150 x 4.6 mm, Ea

Dimensions: 150 x 4.6 mm ID

Order No: 00F-4486-E0

Elution Type: Gradient

Eluent A: Water with 0.1% TFA

Eluent B: Acetonitrile with 0.1% TFA

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	97	3
	2	3	97	3
	3	33	35	65

Flow Rate: 1.2 mL/min

Col. Temp.: ambient


Detection: UV-Vis Abs.-Variable Wave.(UV) @ 214 nm (ambient)

Analyst Note: SecurityGuard[™] Guard Cartridge System extends column lifetime.


- SecurityGuard ULTRA Cartridges, UHPLC WIDEPORE C4 for 4.6mm ID, 3/Pk Part No.: AJ0-8901

- Holder Part No.: AJ0-9000

19883



Products used in this application:



ANALYTES:

1 RNA'ase reduction

