


Paroxetine impurities using Kinetex 5u C18 250x4.6mm**Column:** Kinetex® 5µm C18 100 Å, LC Column 250 x 4.6 mm, Ea**Dimensions:** 250 x 4.6 mm ID**Order No:** 00G-4601-E0**Elution Type:** Gradient**Eluent A:** TFA/Water/THF 5/900/100**Eluent B:** TFA/Acetonitrile/THF 5/900/100

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	90	10
	2	40	20	80

Flow Rate: 1.5 mL/min**Col. Temp.:** ambient**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 295 nm (ambient)

20767

**Products used in this application:****ANALYTES:**

- 1 Impurity A
- 2 Impurity B
- 3 Impurity C
- 4 unidentified
- 5 Paroxetine
- 6 Impurity E

