


uCATs by LCMS using Kinetex 2.6um Polar C18 50x4.6mm**Column:** Kinetex® 2.6 µm Polar C18 100 Å, LC Column 50 x 4.6 mm, Ea**Dimensions:** 50 x 4.6 mm ID**Order No:** 00B-4759-E0**Elution Type:** Gradient**Eluent A:** Water with 0.1% formic acid**Eluent B:** Methanol with 0.1% formic acid

Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	100	0
	2	3	0	100

Flow Rate: 0.7 mL/min**Col. Temp.:** 25 °C**Detection:** Tandem Mass Spec (MS-MS) @ (ambient)**Detector Info:** <a target="_blank"

24172

href="https://sciex.com/products/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIEX<

**Products used in this application:****ANALYTES:**

- 1 Metanephrine
- 2 Normetanephrine
- 3 3-Methoxytyramine



Sample Preparation Details

for HPLC Application ID No.: 24172

uCATs by LCMS using Kinetex 2.6um Polar C18 50x4.6mm

PRODUCT DESCRIPTION:

StrataTM-X-CW Polymeric Weak Cation, microelution 2 mg / well, 96-Well Plates , 1/Pk

Order No.: 8M-S035-4GA

SOLID PHASE EXTRACTION (SPE) PROCEDURE:

Note: The solvent volumes shown below are for a 2 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:

Load:

Wash:

Dry:

3 minutes under high vacuum

Elute:

Final Prep and Analysis:

Pool eluents and dilute with 100 µL of 0.1% Formic acid in water

Inject: 5 µL on HPLC Tandem Mass Spec (MS-MS) @ (ambient)

ANALYTES:	Spiked Conc. (ng/mL)	Log P	pKa	% Rec	%RSC (n=0)
1 Metanephrene	0				
2 Normetanephrene	0				
3 3-Methoxytyramine	0				

Note: This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals.
Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at info@phenomenex.com



Phenomenex products are available worldwide.

www.phenomenex.com.cn

info@phenomenex.com