


**Charge Variants of Trastuzumab using Salt Gradient on bioZen 6  $\mu$ m WCX****Column:** bioZen<sup>TM</sup> 6 $\mu$ m WCX, LC Column 250 x 4.6 mm, Ea**Dimensions:** 250 x 4.6 mm ID**Order No:** 00G-4777-E0**Elution Type:** Gradient**Eluent A:** 20 mM MES, pH 5.6**Eluent B:** 20 mM MES, 300 mM NaCl, pH 5.6

| Gradient Profile: | Step No. | Time (min) | Pct A | Pct B |
|-------------------|----------|------------|-------|-------|
|                   | 1        | 0          | 85    | 15    |
|                   | 2        | 1          | 85    | 15    |
|                   | 3        | 31         | 55    | 45    |
|                   | 4        | 31.1       | 0     | 100   |
|                   | 5        | 34         | 0     | 100   |
|                   | 6        | 35         | 85    | 15    |

**Flow Rate:** 1 mL/min**Col. Temp.:** 25 °C**Detection:** LC/UV (DAD, PDA) @ 280.000000000 nm (nanometers) (ambient)**ANALYTES:****1** Trastuzumab