


**Paraffin Marker Standards on ZB-DHA-PONA 100m x 0.25mm x 0.5um****Column:** Zebtron ZB-DHA-PONA, GC Cap. Column 100m x 0.25mm x 0.5um , Ea**Phase:****Dimensions:** 100 meters x 0.25 mm x 0.5 µm**Order No:** 7MG-G042-17**Oven Profile:** 35°C for 14 min to 60°C @ 1.1°C/min 60°C for 19 min to 280°C @ 2°C/min 280°C for 5 min**Carrier Gas:** Constant Flow Helium, 1.55 mL/min**Injection:** Split 40:1 0.2 µL @ 300°C**Detection:** GC/FID (320°C)**Analyst Note:** The paraffins standard was first diluted 100:1 in pentane and then run with a 40:1 split to avoid column**Products used in this application:**

## Paraffin Marker Standards on ZB-DHA-PONA 100m x 0.25mm x 0.5um

### ANALYTES:

- 1 Pentane
- 2 Hexane
- 3 Heptane
- 4 Octane
- 5 Nonane
- 6 Decane
- 7 Undecane
- 8 Dodecane
- 9 Tridecane
- 10 Tetradecane
- 11 Pentadecane

